

Microelectrophoresis of extracts from single perithecia

J. B. Nasrallah
Cornell University

A. M. Srb.
Cornell University

Follow this and additional works at: <https://newprairiepress.org/fgr>



This work is licensed under a [Creative Commons Attribution-Share Alike 4.0 License](https://creativecommons.org/licenses/by-sa/4.0/).

Recommended Citation

Nasrallah, J. B., and A.M. Srb. (1978) "Microelectrophoresis of extracts from single perithecia," *Fungal Genetics Reports*: Vol. 25, Article 9. <https://doi.org/10.4148/1941-4765.1731>

This Technical Note is brought to you for free and open access by New Prairie Press. It has been accepted for inclusion in Fungal Genetics Reports by an authorized administrator of New Prairie Press. For more information, please contact cads@k-state.edu.

Microelectrophoresis of extracts from single perithecia

Abstract

Microelectrophoresis of extracts from single perithecia

Nasrallah, J.B. and A.M. Srb. **Microelectrophoresis**

of extracts from single perithecia.

Electrophoresis of proteins from a single perithecium can be performed in capillary tubes, according to modification of Grossbach's procedure (1965, *Biochim. Biophys. Acta* 107: 180-182). Glass capillary tubing with 5-6 mm outside diameter and 1/4-3/4 mm inside diameter (Thomas Co.) are cut into 4 cm lengths. The Column Coat

(Canalco) coated tubes are sealed at one end with parafilm. The electrophoretic system and procedure employed are essentially as described for the standard electrophoretic analysis of perithecial extracts (Nasrallah and Srb (1973), *Proc. Nat. Acad. Sci. USA* 70:

1891-1893) with the following modifications: using a 50 μ Hamilton syringe, the capillary tubes are filled with separating gel solution to a depth of .5 cm, and proportionately twice as much stacking gel is used. Pieces of cellulose acetate film (Sepraphore III, Gelman), cut to fit into the capillary tubes, are washed with 0.1 M phosphate buffer, pH 7.0. A single perithecium is quashed onto a cellulose acetate piece. If desired, microscopic analysis of ascus and ascospore morphology can be performed before the cellulose acetate piece is placed on the surface of the stacking gel. A small volume of stacking gel solution is then layered on top of the cellulose acetate piece to prevent diffusion of the proteins back into the electrophoresis buffer. The upper part of the capillary tube is filled with electrophoresis buffer containing Bromphenol Blue as tracking dye. Electrophoresis is run in a standard Canalco disc electrophoresis apparatus, at 1/2 ma per gel. After the tracking dye has migrated 1 cm down the separating gel, the current is stopped and the capillary tubes are immediately submerged in an ice bath to delay band diffusion. The gels are removed from the capillary tubes using a syringe filled with ethylene glycol and provided with a fine needle (No. 27 or 30). The gels are fixed in 10% (w/v) trichloroacetic acid and stained with a 1:10 dilution of a 1% (w/v) aqueous solution of Coomassie Brilliant Blue in 10% (w/v) trichloroacetic acid.

This technique has been successfully applied to: 1) genetic analysis of electrophoretic variants of perithecial proteins, and 2) to their distribution in single perithecia produced by mycelia heterokaryotic for the determinants of the protein variants and for ascus and ascospore shape. (Supported by Grant GM-12953 from the National Institute of General Medical Sciences, USPHS.) - - Section of Botany, Genetic and Development, Cornell University, Ithaca, New York 14853.